



Canine HA (Hyaluronic Acid) ELISA kit

This kit applies competitive inhibition ELISA method for the quantitative detection of HA (Hyaluronic Acid) in canine samples. The microtiter plate provided in this kit has been pre-coated with Dog HA protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Dog HA. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Dog HA in the samples is then determined by comparing the OD of the samples to the standard curve.

| | |
|----------------------|--|
| Catalog No. | 3148647 |
| Size | 96-Wells |
| Product Category | ELISA (Quantitative) |
| Reactivity | Canine |
| Sample | Serum; Plasma; Other Biological Fluids |
| Assay Method | Competitive Inhibition ELISA |
| Assay Duration | 2 hours |
| Sensitivity | 1.81 ng/mL |
| Standard Curve Range | 6.25-400 ng/mL |
| Storage/Stability | -20°C/1 year; 4°C/6 months |