



Human CTXIII (Cross Linked C-Telopeptide Of Type III Collagen) ELISA kit

This kit applies competitive inhibition ELISA method for the quantitative detection of CTXIII (Cross Linked C-Telopeptide Of Type III Collagen) in human samples. The microtiter plate provided in this kit has been pre-coated with Human CTXIII protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CTXIII. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Human CTXIII in the samples is then determined by comparing the OD of the samples to the standard curve.

Catalog No.	3011543
Size	96-Wells
Product Category	ELISA (Quantitative)
Reactivity	Human
Sample	Serum; Plasma; Other Biological Fluids
Assay Method	Competitive Inhibition ELISA
Assay Duration	2 hours
Sensitivity	46.8 pg/mL
Standard Curve Range	156.25-10000 pg/mL
Storage/Stability	-20°C/1 year; 4°C/6 months