



## Human F8 (Coagulation Factor VIII) ELISA kit

---

This kit applies sandwich ELISA method for the quantitative detection of F8 (Coagulation Factor VIII) in human samples. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human F8. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human F8. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human F8, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$ 10nm. The concentration of Human F8 in the samples is then determined by comparing the OD of the samples to the standard curve.

Catalog No.	3012249
Size	96-Wells
Product Category	ELISA (Quantitative)
Reactivity	Human
Sample	Plasma
Assay Method	Sandwich ELISA
Assay Duration	3.5 hours
Sensitivity	30 pg/mL
Standard Curve Range	78.13-5000 pg/mL
Storage/Stability	-20°C/1 year; 4°C/6 months