



## Human GC (Glucagon) ELISA kit

---

This kit applies sandwich ELISA method for the quantitative detection of GC (Glucagon) in human samples. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human GC. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human GC. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human GC, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$ 10nm. The concentration of Human GC in the samples is then determined by comparing the OD of the samples to the standard curve.

Catalog No.	3012159
Size	96-Wells
Product Category	ELISA (Quantitative)
Reactivity	Human
Sample	Plasma
Assay Method	Sandwich ELISA
Assay Duration	3.5 hours
Sensitivity	7.3 pg/mL
Standard Curve Range	15.63-1000 pg/mL
Storage/Stability	-20°C/1 year; 4°C/6 months