

Human Muscarinic Acetylcholine Receptor (M-AChR) ELISA kit

This kit applies sandwich ELISA method for the quantitative detection of Muscarinic Acetylcholine Receptor (M-AChR) in human samples. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human M-AChR. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human M-AChR. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human M-AChR, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ±10nm. The concentration of Human M-AChR in the samples is then determined by comparing the OD of the samples to the standard curve.

| Catalog No. | 3018532 |
|----------------------|--|
| Size | 96-Wells |
| Product Category | ELISA (Quantitative) |
| Reactivity | Human |
| Sample | Serum; Plasma; Other Biological Fluids |
| Assay Method | Sandwich ELISA |
| Assay Duration | 3.5 hours |
| Sensitivity | 0.135 ng/mL |
| Standard Curve Range | 0.32-20 ng/mL |
| Storage/Stability | -20°C/1 year; 4°C/6 months |

www.realgenelabs.com

For Research Use only