

## Porcine ADST (Androstenone) ELISA kit

---

This kit applies competitive inhibition ELISA method for the quantitative detection of ADST (Androstenone) in porcine samples. The microtiter plate provided in this kit has been pre-coated with Pig ADST protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Pig ADST. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$ 10nm. The concentration of Pig ADST in the samples is then determined by comparing the OD of the samples to the standard curve.

Catalog No.	3218586
Size	96-Wells
Product Category	ELISA (Quantitative)
Reactivity	Porcine
Sample	Serum; Plasma; Other Biological Fluids
Assay Method	Competitive Inhibition ELISA
Assay Duration	2 hours
Sensitivity	10.19 nmol/ml
Standard Curve Range	31.25-2000 nmol/ml
Storage/Stability	-20°C/1 year; 4°C/6 months