

## Sheep vWF (Von Willebrand Factor) ELISA kit

---

This kit applies competitive inhibition ELISA method for the quantitative detection of vWF (Von Willebrand Factor) in sheep samples. The microtiter plate provided in this kit has been pre-coated with Sheep vWF protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Sheep vWF. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$ 10nm. The concentration of Sheep vWF in the samples is then determined by comparing the OD of the samples to the standard curve.

Catalog No.	3239355
Size	96-Wells
Product Category	ELISA (Quantitative)
Reactivity	Sheep
Sample	Plasma
Assay Method	Competitive Inhibition ELISA
Assay Duration	2 hours
Sensitivity	4.11 ng/mL
Standard Curve Range	12.5-800 ng/mL
Storage/Stability	-20°C/1 year; 4°C/6 months