

Malondialdehyde (MDA) Rapid ELISA kit

This Rapid ELISA kit applies Competitive Inhibition enzyme immunoassay technique for the quantitative detection. The micro ELISA plate provided in this kit has been pre-coated with Malondialdehyde (MDA) protein. Standards or samples are added to the appropriate micro ELISA plate wells then with a biotin-conjugated antibody specific to Malondialdehyde (MDA) . Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Malondialdehyde (MDA) in the samples is then determined by comparing the OD of the samples to the standard curve.

Catalog No.	3510165
Size	96-Well Tests
Product Category	Rapid ELISA
Reactivity	Various
Sample	Serum; Plasma; Other Biological Fluids
Assay Method	Competitive Inhibition ELISA
Assay Duration	45 min
Sensitivity	9.15 ng/mL
Standard Curve Range	31.25-2000 ng/mL
Storage/Stability	-20°C/1 year; 4°C/6 months